

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/24/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14260

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:17
INTERNAL STANDARD	VERIFIED	00:18
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:18
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	33.92c	
REFERENCE STANDARD	0.09	00:19
BLANK TEST	0.00	00:20
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	00:21
BLANK TEST	0.00	00:21



Sgt Ray Brown WCSO #16
02-24-15