

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/10/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:18
INTERNAL STANDARD	VERIFIED	03:18
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:18
BLANK TEST	0.00	03:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:19
BLANK TEST	0.00	03:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:21
BLANK TEST	0.00	03:22

Sgt R J Bartlett

WCSO #162 03-10-15

