

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/24/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14260

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:29
INTERNAL STANDARD	VERIFIED	00:29
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:29
BLANK TEST	0.00	00:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:32
BLANK TEST	0.00	00:33



Sgt R.J. Bartlett weso #162
03-24-15