

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 04/18/2015

OPERATOR NAME:  
JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:18
INTERNAL STANDARD	VERIFIED	15:18
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	15:18
BLANK TEST	0.00	15:19
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	15:20
BLANK TEST	0.00	15:21
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	15:21
BLANK TEST	0.00	15:22

*J. Lopez #299*

