

SOLUTION CHANGE

-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 105208  
 -----

Date: 05/04/2015

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:51
INTERNAL STANDARD	VERIFIED	16:51
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	16:51
BLANK TEST	0.00	16:52
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	16:53
BLANK TEST	0.00	16:54
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	16:54
BLANK TEST	0.00	16:55

*SGS Bk of Results WCSO #162*  
 05-04-15

