

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/13/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:37
INTERNAL STANDARD	VERIFIED	14:37
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	14:37
BLANK TEST	0.00	14:38
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	14:39
BLANK TEST	0.00	14:40
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	14:40
BLANK TEST	0.00	14:41



SGS Dept of Justice WGS 5012 12

05-13-15