

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 06/08/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15070

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:57
INTERNAL STANDARD	VERIFIED	02:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:58
BLANK TEST	0.00	02:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:00
BLANK TEST	0.00	03:01



Sgt Ralph Bartlett W/50 Sheriff #
 162
 06-08-15