

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 06/30/2015

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 15070

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:47
INTERNAL STANDARD	VERIFIED	02:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:48
BLANK TEST	0.00	02:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:49
BLANK TEST	0.00	02:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:51
BLANK TEST	0.00	02:52



SGS RJA Bartlett WCSO#162  
06-30-15