

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/28/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15180

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:54
INTERNAL STANDARD	VERIFIED	00:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:54
BLANK TEST	0.00	00:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:56
BLANK TEST	0.00	00:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:57
BLANK TEST	0.00	00:58

*SOS Blk / Results w/ISO #16c
09-28-15*

