

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/02/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15180

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:25
INTERNAL STANDARD	VERIFIED	18:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:25
BLANK TEST	0.00	18:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:27
BLANK TEST	0.00	18:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:28
BLANK TEST	0.00	18:29

*SW Best / Rueda WLSO#122
11/2/15*

