

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/15/2015

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15180

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:57
INTERNAL STANDARD	VERIFIED	09:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:58
BLANK TEST	0.00	09:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	10:00
BLANK TEST	0.00	10:01



Charles B. Geisler #298
11/15/15 1002