

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 105208  
 -----

Date: 12/14/2015

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 15210

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:31
INTERNAL STANDARD	VERIFIED	15:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:31
BLANK TEST	0.00	15:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:32
BLANK TEST	0.00	15:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:34
BLANK TEST	0.00	15:35



*Sgt. A. J. Bartlett was #162  
 12-14-15*