

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/29/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15210

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:34
INTERNAL STANDARD	VERIFIED	02:34
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:34
BLANK TEST	0.00	02:35
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:36
BLANK TEST	0.00	02:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:37
BLANK TEST	0.00	02:38



SGS Rep Bartlett Waco #162
12/29/15