

SOLUTION CHANGE




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106606  
 -----

Date: 03/05/2015

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 14260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:17
INTERNAL STANDARD	VERIFIED	03:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:18
BLANK TEST	0.00	03:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:19
BLANK TEST	0.00	03:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:20
BLANK TEST	0.00	03:21



Sgt *RJ Bartlett*  
 WCSO #162 03/05/15