

SUPERVISOR TEST




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 03/24/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 14260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:14
INTERNAL STANDARD	VERIFIED	03:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:15
BLANK TEST	0.00	03:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:16
BLANK TEST	0.00	03:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:17
BLANK TEST	0.00	03:18



SGS Rly Barrett WOSO #16c
 03-24-15