

SUPERVISOR TEST




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 04/07/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:18
INTERNAL STANDARD	VERIFIED	00:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:18
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:20
BLANK TEST	0.00	00:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:21
BLANK TEST	0.00	00:22



565 Sgt J Bartlett Waco #162
 04-07-15