

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 04/18/2015

OPERATOR NAME:  
JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:38
INTERNAL STANDARD	VERIFIED	16:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:38
BLANK TEST	0.00	16:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:39
BLANK TEST	0.00	16:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:41
BLANK TEST	0.00	16:42

*J. M. Lopez 299*

