

SOLUTION CHANGE




WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106606

Date: 05/04/2015

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:36
INTERNAL STANDARD	VERIFIED	22:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:37
BLANK TEST	0.00	22:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:38
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:39
BLANK TEST	0.00	22:40



*Sus Rly Bartlett #162*

*05-04-15*