

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 06/02/2015

OPERATOR NAME:
RALPH J BARTLET
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15070

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:40
INTERNAL STANDARD	VERIFIED	02:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:41
BLANK TEST	0.00	02:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:42
BLANK TEST	0.00	02:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:43
BLANK TEST	0.00	02:44



565 Rly ~~Result~~ #162
06-02-15