

SUPERVISOR TEST


WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 06/22/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15070

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:53
INTERNAL STANDARD	VERIFIED	02:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:54
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:56
BLANK TEST	0.00	02:57



See Rly Report was 07/16
06-22-15