

SUPERVISOR TEST




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 07/06/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15070

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:56
INTERNAL STANDARD	VERIFIED	00:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:57
BLANK TEST	0.00	00:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:58
BLANK TEST	0.00	00:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:59
BLANK TEST	0.00	01:00



505 Rly Bureau w/ SU #162
 07-06-15