

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/25/2015

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15070

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:46
INTERNAL STANDARD	VERIFIED	11:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:46
BLANK TEST	0.00	11:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:48
BLANK TEST	0.00	11:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:49
BLANK TEST	0.00	11:50

J. Lopez 209

