

SOLUTION CHANGE


WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 08/01/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15140

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:16
INTERNAL STANDARD	VERIFIED	00:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:16
BLANK TEST	0.00	00:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:18
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:19
BLANK TEST	0.00	00:20



SGS RJA Bartlett WCSO #162
08-01-15