

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 08/12/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:14
INTERNAL STANDARD	VERIFIED	03:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:15
BLANK TEST	0.00	03:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:16
BLANK TEST	0.00	03:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:17
BLANK TEST	0.00	03:18



565 Page 1 Bartlett wco #162
08-12-15