

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 08/17/2015

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15140

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:34
INTERNAL STANDARD	VERIFIED	19:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:34
BLANK TEST	0.00	19:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:36
BLANK TEST	0.00	19:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:37
BLANK TEST	0.00	19:38

Ma. J. Lopez

