

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 08/31/2015

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 15140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:52
INTERNAL STANDARD	VERIFIED	03:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:53
BLANK TEST	0.00	03:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:54
BLANK TEST	0.00	03:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:55
BLANK TEST	0.00	03:56



SGS Rly Jones wesu #102

08-31-15