

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 09/08/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15140

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:00
INTERNAL STANDARD	VERIFIED	04:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:01
BLANK TEST	0.00	04:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:02
BLANK TEST	0.00	04:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:03
BLANK TEST	0.00	04:04

SGK Rly Russo WCSO #162
09-08-15

