

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 09/28/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222Q
Operator Agency/Dept: WARREN COUNTY
SHERIFF'S OFFICE
Reference Standard #: 15180

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:49
INTERNAL STANDARD	VERIFIED	02:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:49
BLANK TEST	0.00	02:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:51
BLANK TEST	0.00	02:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:52
BLANK TEST	0.00	02:53



SG *RJ Bartlett* WCSO #162
09-28-15