

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 10/12/2015

OPERATOR NAME:  
KENNETH W SMITH  
Permit #: 35604  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 15180

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:41
INTERNAL STANDARD	VERIFIED	03:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:42
BLANK TEST	0.00	03:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:43
BLANK TEST	0.00	03:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:44
BLANK TEST	0.00	03:45



*PO. [Signature] # 301*

*10/12/15*