

SUPERVISOR TEST




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 10/20/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15180

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:27
INTERNAL STANDARD	VERIFIED	03:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:27
BLANK TEST	0.00	03:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:28
BLANK TEST	0.00	03:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:30
BLANK TEST	0.00	03:31



SGS Rjt 1 *Russell* WCSO #162
 10-20-15