

SUPERVISOR TEST


WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 12/14/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15210

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:50
INTERNAL STANDARD	VERIFIED	16:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:51
BLANK TEST	0.00	16:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:52
BLANK TEST	0.00	16:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:53
BLANK TEST	0.00	16:54



SUS Rly Thru waso #162
12-14-15