

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 01/05/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 14230

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:20
INTERNAL STANDARD	VERIFIED	01:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:22
BLANK TEST	0.00	01:22
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:23
BLANK TEST	0.00	01:24



SGS Br. / Bartlett
 WCSO # 162 01-0575