

SUPERVISOR TEST




 WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 01/13/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 14230

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:43
INTERNAL STANDARD	VERIFIED	00:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:43
BLANK TEST	0.00	00:44
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:44
BLANK TEST	0.00	00:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:46
BLANK TEST	0.00	00:46



SGS *RJ Bartlett* WCSO #162
 01-13-15