

SUPERVISOR TEST



WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 01/27/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 14230

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:01
INTERNAL STANDARD	VERIFIED	02:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:01
BLANK TEST	0.00	02:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:02
BLANK TEST	0.00	02:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:03
BLANK TEST	0.00	02:04



SG: *RJ Bartlett* W.C. SO #162
 01-27-15