

SOLUTION CHANGE

WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/02/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 14260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:53
INTERNAL STANDARD	VERIFIED	01:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:54
BLANK TEST	0.00	01:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:55
BLANK TEST	0.00	01:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:56
BLANK TEST	0.00	01:57

SGS By Bartlett WCR #162
 02-02-15

