

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/24/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 14260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:43
INTERNAL STANDARD	VERIFIED	01:43
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:43
BLANK TEST	0.00	01:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:45
BLANK TEST	0.00	01:45
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:46
BLANK TEST	0.00	01:47



SGT *Ralph Bartlett* WISO #162
 02-24-15