

SOLUTION CHANGE




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 03/05/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 14260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:26
INTERNAL STANDARD	VERIFIED	01:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:27
BLANK TEST	0.00	01:28
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:28
BLANK TEST	0.00	01:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:29
BLANK TEST	0.00	01:30



SCS *RJ Bartlett*
 WCSO SHERRIF #166 03-05-15