

SUPERVISOR TEST




WARREN COUNTY  
 SHERRIF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 03/24/2015

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 14260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:07
INTERNAL STANDARD	VERIFIED	01:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:07
BLANK TEST	0.00	01:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:09
BLANK TEST	0.00	01:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:10
BLANK TEST	0.00	01:11



SGS RJB Bartlett Wes #162  
 03-24-15