

SOLUTION CHANGE




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 03/30/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:01
INTERNAL STANDARD	VERIFIED	01:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:01
BLANK TEST	0.00	01:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:02
BLANK TEST	0.00	01:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:04
BLANK TEST	0.00	01:05



505 *Ralph Bartlett* WCO #142
 03-30-15