

SUPERVISOR TEST




WARREN COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 04/13/2015

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:52
INTERNAL STANDARD	VERIFIED	00:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:53
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:54
BLANK TEST	0.00	00:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:55
BLANK TEST	0.00	00:56



*Sgt R.J. Bartlett* Permit #162  
 04-13-15