

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/18/2015

OPERATOR NAME:  
JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:41
INTERNAL STANDARD	VERIFIED	15:42
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	15:42
BLANK TEST	0.00	15:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	15:43
BLANK TEST	0.00	15:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:44
BLANK TEST	0.00	15:45

*J.M. Lopez 299*

