

SUPERVISOR TEST

WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/27/2015

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:14
INTERNAL STANDARD	VERIFIED	09:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:15
BLANK TEST	0.00	09:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:16
BLANK TEST	0.00	09:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:17
BLANK TEST	0.00	09:18

