

SOLUTION CHANGE




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/04/2015

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:27
INTERNAL STANDARD	VERIFIED	17:27
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	17:28
BLANK TEST	0.00	17:29
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	17:29
BLANK TEST	0.00	17:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:30
BLANK TEST	0.00	17:31



505 *Ralph Bartlett* #162 05/04/15