

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 05/13/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:08
INTERNAL STANDARD	VERIFIED	14:08
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	14:09
BLANK TEST	0.00	14:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:10
BLANK TEST	0.00	14:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:11
BLANK TEST	0.00	14:12



SES RJB
 WCSO #162 05/13-15