

SUPERVISOR TEST




 WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 05/18/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15010

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:27
INTERNAL STANDARD	VERIFIED	15:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:27
BLANK TEST	0.00	15:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:29
BLANK TEST	0.00	15:30
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	15:30
BLANK TEST	0.00	15:31



SOB BSL / Printed # 162
 05-18-15