

SOLUTION CHANGE




WARREN COUNTY  
 SHERRIF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 06/30/2015

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 15070

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:30
INTERNAL STANDARD	VERIFIED	01:30
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:31
BLANK TEST	0.00	01:31
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:32
BLANK TEST	0.00	01:33
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:33
BLANK TEST	0.00	01:34



*Sus Rly R. Bartlett*

WC90 # 162

06-30-15