

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/19/2015

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15070

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:17
INTERNAL STANDARD	VERIFIED	00:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:18
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:19
BLANK TEST	0.00	00:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:20
BLANK TEST	0.00	00:21



SGS Rep of Bartlett WOSO #162
07-19-15