

UPERVISOR TEST

WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/25/2015

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 15070

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:25
INTERNAL STANDARD	VERIFIED	19:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:26
BLANK TEST	0.00	19:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:27
BLANK TEST	0.00	19:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:28
BLANK TEST	0.00	19:29

J. Lopez 7/29/15

