

SOLUTION CHANGE

WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 08/17/2015

OPERATOR NAME:
 JOSHUA M LOPEZ
 Permit #: 34728
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15140

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:53
INTERNAL STANDARD	VERIFIED	18:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:53
BLANK TEST	0.00	18:54
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	18:55
BLANK TEST	0.00	18:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:56
BLANK TEST	0.00	18:57

J. M. Lopez 289

