

SUPERVISOR TEST

WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 08/25/2015

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 15140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:51
INTERNAL STANDARD	VERIFIED	02:51
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:52
BLANK TEST	0.00	02:53
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:54
BLANK TEST	0.00	02:55

Sgt R.J. Bartlett WRS #162
08-25-15

